


OEHL
GROUP LTD

Understanding your environment

Full Service Range



contact Tracey Scott on 0121 359 5361
or email enquiries@oehgroup.co.uk
www.oehgroup.co.uk

**MEASUREMENT OF ENVIRONMENTAL EMISSIONS
FROM
METAL COATING OPERATIONS**

at

**LONDON TAXIS INTERNATIONAL
HOLYHEAD ROAD
COVENTRY
WARWICKSHIRE
CV5 8JJ**

REPORT NO:	OEH/31689a/STAK/SL42	CLIENT REF:	Purchase Order No: 5709
DATE OF VISIT:	19 November, 2003	CONTACT ON SITE:	Chris Cuffin
DATE OF REPORT:	10 January, 2005	DISK REFERENCE:	N:\Consultants\$\Air Quality\Andy Barnes\Jobs 2003-2004\London Taxis International For reissue\OEH31689a CS F.doc 10/01/2005 13:00

DATA PROTECTION ACT REGISTRATION NO: B0479 03 4

CONFIDENTIALITY UNDERTAKING

We undertake that we will not knowingly make use or disclose any confidential information or photographs relating to your business which may have come to our knowledge or attention as a result of our visit on site or otherwise as a result of the work carried out by us in connection with the preparation of this report. If you have any queries or comments regarding this report, please contact the Customer Services, OEH Group Limited Tel: 0121 359 5361.

TABLE OF CONTENTS

EXECUTIVE SUMMARY	3
1 INTRODUCTION	4
1.1 PURPOSE OF SURVEY	4
1.2 TERMS OF REFERENCE	4
1.3 PLANT CONDITIONS	4
2 PROCESS DESCRIPTION	4
3 METHODS	5
3.1 STACK SAMPLING	5
3.1.1 Stack Velocity & Temperature Measurements	5
3.1.2 Total Particulate Matter	5
3.1.3 Volatile Organic Compounds (VOC)	5
3.1.4 Isocyanates	5
3.2 ANALYSIS	6
3.2.1 Techniques & Detection Limits	6
3.2.2 Accreditation	6
4 PRESENTATION OF RESULTS	6
5 DISCUSSION	7
5.1 VOLATILE ORGANIC COMPOUNDS	7
5.2 PARTICULATE MATTER	7
5.3 ISOCYANATES	7
5.4 MONITORING DEVIATIONS	8
6 CONCLUSIONS	8
7 APPENDICES	8
APPENDIX I: DETAILED PARTICULATE & FLOWRATE RESULTS TABLES	8
APPENDIX II: VOC PROFILING DATA	8
APPENDIX III: CALIBRATION CERTIFICATES	8

EXECUTIVE SUMMARY

Date Of Test & Test Areas Emissions sampling from the Combi Booth Stack (Main Paint Plant) was conducted on 19th November 2003.



Test Conditions The process was being operated under normal conditions throughout the sampling period.
Further details are given below in section 2 of this report.

Compliance In the absence of any details regarding the authorisation for the site, the emissions have been compared to the limits set out in Process Guidance Note PG6/34(2004) - Respraying of Road Vehicles.

General Observations The stack serving the clearcoat section of the main booth was also scheduled for sampling on the test date; however no safe access to the sampling position was available so this position could not be sampled.

Surveyed and reported by:

Verified by:



P.R. Bill Roberts
Environmental Technician

Andy Barnes *BSc (Hons)*
Environmental Scientist

for and on behalf of OEH Group Limited

If you have any queries or comments regarding this report, please contact Customer Services, OEH Group Ltd. Tel: 0121 359 5361.

1 INTRODUCTION

1.1 Purpose of Survey

The aim of the survey described in this report was to verify compliance with the requirements of the relevant Process Guidance Note, PG6/34(2004) – Respraying of Road Vehicles.

1.2 Terms of Reference

London Taxis International, Holyhead Road, Coventry, Warwickshire, CV5 8JJ, has commissioned OEH Group Limited to carry out the work described in this report. Monitoring was carried out on 19th November 2003, by Bill Roberts, at the request of Chris Cuffin.

The work was carried out in accordance with OEH Proposal ref: EFH-9999, dated 9th May 2003 and with the client's instructions, as set out in Purchase Order 5709.

OEH Group is accredited under ISO-9002 for the provision of health, safety and environmental consultancy services. The work described in this report was carried out in accord with our ISO-9000 Standard Operating Procedures and Level III: Consultancy Work Instructions.

The field sampling and interpretations made in this report are not covered by the scope of OEH's accreditation under UKAS.

1.3 Plant conditions

Production schedules on the dates of the survey were described as normal. Thus, the data reported herein must be considered typical and representative of the environmental levels experienced during normal daily workloads on this site.

2 PROCESS DESCRIPTION

LTI undertake the manufacture of traditional style London Taxis. The production facility utilises a large number of processes. This survey deals only with emissions from the main paint plant combi spraybooth.

Small panels are sprayed in the booth, during this operation the extraction system from the booth is set to discharge to atmosphere.

Once spraying is completed, the exhaust air is heated and recirculated back into the booth for drying, at this stage there is no emission to atmosphere.

Once panels are dry, the heating is switched off and the extraction is again set to discharge to atmosphere.

During the test period the following timings were noted:

10:06 to 11:20 – Little or no activity, extraction exhausting to atmosphere.

11:20 to 11:52 – Spraying takes place, extraction exhausting to atmosphere.

11:52 to 12:30 – Stoving taking place, extracted air heated and recirculated – no flow in emissions stack.

12:30 to 13:45 – Items cooling, extraction exhausting to atmosphere.

3 METHODS

3.1 Stack Sampling

3.1.1 Stack Velocity & Temperature Measurements

Stack velocity was investigated using an ellipsoidal nosed pitot tube coupled to an electronic manometer. Temperature measurements were taken using a K-type thermocouple connected to an electronic thermometer.

The manometer and thermometer are subject to regular calibration by a UKAS accredited test house using NPL (National Physics Laboratory) traceable standards.

3.1.2 Total Particulate Matter

Periodic extractive sampling for total particulate matter was conducted using a Stackmite 9096 sampling train, incorporating an in-stack, 47mm GFA filter. Within the limitations of the stack and field conditions, the sampling protocol was in accordance with the main procedural requirements of BS EN 13284-1:2002. The sampling train was set up and checked for leaks before commencement of the survey and between each sample, all leak rates were below 2% of the maximum sampling flowrate. The Stackmite unit is calibrated annually and is traceable to NPL standards. Calibration dated 31st March 2003.

Once the fan had stopped, sampling was continued at the rate calculated during spraying. This means that the reported emissions may be an overestimate of the true levels.

BSEN 13284-1:2002 was used instead of BS ISO 9096:2003 as requested in PG6/34(2004) as the former standard is specifically designed for low level concentrations expected in the emissions.

BS ISO 9096:2003 is only designed for use in the range 20 – 1000 mg.m⁻³ and BSEN 13284-1:2002 is for use at any concentration below 50 mg.m⁻³.

3.1.3 Volatile Organic Compounds (VOC)

Continuous extractive sampling for VOCs was conducted using a Research Engineers Flame Ionisation Detector. The instrument was calibrated on site using standard methane span gas traceable to an NPL standard. Results are expressed as methane equivalent values. The sampling protocol was in accordance with the main procedural requirements of BS EN 13526:2002.

Continuous extractive sampling was backed up by periodic extractive sampling for VOCs using a calibrated pump connected to charcoal adsorption tubes. The method is based on, and intended to satisfy the main procedural requirements of BS EN 13649:2002. The results of this tube sampling were used to provide a correction factor for the FID sampling.

3.1.4 Isocyanates

Periodic extractive sampling for Isocyanates was conducted using a calibrated pump connected to an impinger sampling train containing a solution of 1-(2-methoxyphenyl)piperazine. The method is based on, and intended to satisfy the main procedural requirements of BS ISO 16702.

3.2 Analysis

3.2.1 Techniques & Detection Limits

Analyte	Analysis Technique	Detection Limit	Analytical Precision, %	Method Reference
TPM	Gravimetric	20 µg	1	LSOP 202
Isocyanates	HPLC	0.02 µg NCO	10	LSOP 502
Continuous VOC	Flame Ionisation Detector	0.2 mg.m ⁻³ as carbon	5	BS EN 13526
Periodic VOC	Gas Chromatography	2 µg as carbon	5	Variation on LSOP 402

3.2.2 Accreditation

Service Category	ISO-9002	UKAS ¹
Consultancy	Yes	No
Analysis		
- Dusts (air filter samples); Lab Method LSOP 202, based on MDHS14 (latest issue)	Yes	Yes
- Solvents (B, T, X 111-T, TCE, PERC); Lab Method LSOP402, based on Various NIOSH	Yes	Yes
- Solvents (all other species); Based on Various NIOSH	Yes	No
- Isocyanates (impinger samples); Lab Method LSOP 502, based on MDHS25 (latest issue)	Yes	No
¹ UKAS lab number 1821		
<i>Stack sampling team is a member of the Source Testing Association</i>		

4 PRESENTATION OF RESULTS

The following table gives summary details of the mean emission concentrations measured for all parameters from all positions.

Sampling Position	Mean Particulate Emission (mg.m ⁻³)	Mean Isocyanate Emission (mg.m ⁻³)	Mean VOC Emission (mgC.m ⁻³)
Main Paint Plant Combi Booth	4.2 (11:30 – 11:55)	<0.001 (11:30 – 11:52)	4.9 (10:07 – 13:45)

Results reported at Standard Conditions of 273K and 101.3kPa, no correction for water vapour content.

Detailed results for all sampling positions are included in the Appendices of this report as follows:

Appendix I lists in tabular form further details of the particulate results, including additional data from the pitot traverses, along with filter weight details and sampling parameters.

The recorded data for VOC concentrations is presented in both graphical and tabular form in Appendix II. This includes the 2-minute and 15-minute average data for VOCs.

5 DISCUSSION

The coating processes are covered by the Secretary of States Guidance Note, PG6/34(2004) – Respraying of Road Vehicles, which gives the following limits

Parameter	Emissions Limit
Total Particulate Matter	10 mg.m ⁻³
Isocyanates (expressed as total NCO group)	non stated (0.1 mg.m ⁻³ is typical limit in other guidance notes)
Volatile Organic Compounds (as total carbon excluding particulate matter)	50 mg.m ⁻³ as: 2 minute mean averages for spraying operations. 15 minute averages for baking operations.

5.1 Volatile Organic Compounds

The average measured emission over the entire test period was 4.9 mg.m⁻³. This result is representative of the entire process.

The average over the spraying period only (11:20 – 11:52) was 12.7 mg.m⁻³. This result is representative of worst case emissions.

The highest two minute average over the spraying period was 51.9 mg.m⁻³.

The highest fifteen minute average over the bake period was 6.4 mg.m⁻³.

5.2 Particulate Matter

The measured particulate emission concentration was 4.2 mg.m⁻³ and was therefore well below the 10 mg.m⁻³ emissions limit.

The sample was taken at the point of most likely emissions when spraying was occurring; this was therefore a worst case emission level.

5.3 Isocyanates

Isocyanate emissions were less than the 0.001 mg.m⁻³ analytical limit of detection and as such were well below the typical emissions limit of 0.1 mg.m⁻³.

The sample was taken at the point of most likely emissions when spraying was occurring; this was therefore a worst case emission level.

5.4 Monitoring Deviations

Emission Point Reference	Substance Deviations	Monitoring Deviations	Other Relevant Issues
Main Paint Plant Combi Booth	None	Yes - See Below	None

The sampling plane does not meet the flow requirements of BS EN 13284-1:2002 – Section 5.2, for the following reasons:

Ratio of highest to lowest gas velocity exceeded the allowed 3:1 level at 3.66:1; however there was no safe alternative position available.

There were sampling deviations from BS EN 13284-1:2002 as follows:

Only 1 sample could be performed due to the short amount of spraying taking place in the booth, the test was for 25 minutes only.

Only two sampling positions were used on one sampling line, this is because the second sampling line could not be safely accessed.

The isokinetic rate was just outside the allowed criterion of -5 to +15% at -5.6%.

6 CONCLUSIONS

From the data reported it can be seen that the process does not demonstrate compliance with the emission limits set down in the relevant Process Guidance Notes under normal and typical workloads. One two minute VOC average from the spray part of the cycle marginally exceeded the emissions limit.

7 APPENDICES

Appendix I: Detailed Particulate & Flowrate Results Tables

Appendix II: VOC Profiling Data

Appendix III: Calibration Certificates

APPENDIX I
DETAILED PARTICULATE & FLOWRATE RESULTS TABLES

Plant Type	Combi Booth	Stack Area (m ²)	0.454
Job Number	OEH 31689	Meter Temp (C)	30
Client Name	London Taxis	Stack Diameter (cm)	76
Date	19th November 2003	Pitot Factor	1.00
		Pitot Factor (sqrt)	1.00
		Stack Pressure (Pa)	50
		Ambient Pressure (kPa)	101.3
		Nozzle Size (mm)	6.00

PITOT SURVEY

Traverse Point	A1	A2	A3	A4	A5	A6	A7	A8	A9	A10
Distance From Near Wall (D)	0.065	0.150	0.250	0.350	0.450	0.550	0.650	0.750	0.850	0.935
Pitot Reading (Pa)	150	130	50	22	18	12	11	20	38	60
Temperature (°C)	31	31	31	31	31	31	31	31	31	31
Duct Velocity (m/s)	16.1	15.0	9.3	6.2	5.6	4.5	4.4	5.9	8.1	10.2

Absolute Mean Duct Velocity (m/s) 8.5
Absolute Flow Rate (m³/hr) 13903
Normalised Flow Rate (Nm³/hr) 12492

Sampling Run 1 Time: 11:30 - 11:55

Sampling Point	A3	A8	Initial Meter Reading (l)	691150
Sampling Rate (l/min)	15	10	Final Meter Reading (l)	691445
Sampling Duration (mins)	12.5	12.5	Volume Sampled (l)	295
Filter No	8110	8110	Isokineticity Error (%)	-5.6
Volume Sampled (m ³)	Meter 0.295	Expected 0.313	<i>(Maximum Allowed Error = -5 to +15%)</i>	
Corrected Volume =	0.27 Nm ³ (at NTP)			

FILTER WEIGHTS


Test Number	Filter No	Pre-Weight (mg)	Post-Weight (mg)	Gain (mg)
1	8110	56.20	57.32	1.12

TEST RESULTS

	Test 1	Mean
Particulate Concentration(mg/Nm ³)	4.2	4.2
Mass Emission (g/hr)	53	53

APPENDIX II
VOC PROFILING DATA

VOC Profiling Data - London Taxis International - Main Paint Plant Combi Booth - 19/11/03




Sampling Time [hh:mm]


----- Emission Concentration

----- Emission Limit

2 Minute Average Values - VOC Data During Set up and Spraying - London Taxis International -
Main Paint Plant Combi Booth - 19/11/03



15 Minute Average Values - VOC Data During Bake & Cooling - London Taxis International -
 Main Paint Plant Combi Booth - 19/11/03



Job Ref: OEH 31689
Client Name: LTI
Location: Combi Booth
Date: 19-Nov-03
Scientist: BR

Technical Details
Instrument Type: FID
Calib Gas: Methane
% C: 75
Sample Number: SA 5376
Instrument Range: 440
Emission Limit: 50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
10:06:59	0.2	31	0.53
10:07:19	0.2	31	0.53
10:07:39	0.4	31	1.14
10:07:59	0.2	31	0.53
10:08:19	0.4	31	1.14
10:08:39	0.4	31	1.14
10:08:59	0.4	31	1.14
10:09:19	0.4	31	1.14
10:09:39	0.4	31	1.14
10:09:59	0.5	31	1.74
10:10:19	0.5	31	1.74
10:10:39	0.5	31	1.74
10:10:59	0.4	31	1.14
10:11:19	0.4	31	1.14
10:11:39	0.5	31	1.74
10:11:59	0.4	31	1.14
10:12:19	0.5	31	1.74
10:12:39	0.5	31	1.74
10:12:59	0.5	31	1.74
10:13:19	0.4	31	1.14
10:13:39	0.4	31	1.14
10:13:59	0.4	31	1.14
10:14:19	0.4	31	1.14
10:14:39	0.4	31	1.14
10:14:59	0.4	31	1.14
10:15:19	0.4	31	1.14
10:15:39	0.4	31	1.14
10:15:59	0.4	31	1.14
10:16:19	0.4	31	1.14
10:16:39	0.4	31	1.14
10:16:59	0.4	31	1.14
10:17:19	0.4	31	1.14
10:17:39	0.4	31	1.14
10:17:59	0.4	31	1.14
10:18:19	0.4	31	1.14
10:18:39	0.2	31	0.53
10:18:59	0.4	31	1.14
10:19:19	0.4	31	1.14
10:19:39	0.4	31	1.14
10:19:59	0.4	31	1.14
10:20:19	0.4	31	1.14
10:20:39	0.4	31	1.14
10:20:59	0.4	31	1.14
10:21:19	0.4	31	1.14

Job Ref:

OEH 31689

Technical Details

Client Name:

LTI

Instrument Type

FID

Location:

Combi Booth

Calib Gas

Methane

Date:

19-Nov-03

% C:

75

Scientist:

BR

Sample Number:

SA 5376

Instrument Range:

440

Emission Limit:

50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
10:21:39	0.4	31	1.14
10:21:59	0.4	31	1.14
10:22:19	0.5	31	1.74
10:22:39	0.5	31	1.74
10:22:59	0.5	31	1.74
10:23:19	0.5	31	1.74
10:23:39	0.5	31	1.74
10:23:59	0.5	31	1.74
10:24:19	1.1	31	3.41
10:24:39	0.5	31	1.74
10:24:59	0.7	31	2.27
10:25:19	0.7	31	2.27
10:25:39	0.7	31	2.27
10:25:59	0.7	31	2.27
10:26:19	0.7	31	2.27
10:26:39	0.7	31	2.27
10:26:59	0.7	31	2.27
10:27:19	0.9	31	2.88
10:27:39	1.1	31	3.41
10:27:59	1.1	31	3.41
10:28:19	1.2	31	4.01
10:28:39	1.2	31	4.01
10:28:59	1.1	31	3.41
10:29:19	0.9	31	2.88
10:29:39	1.2	31	4.01
10:29:59	1.1	31	3.41
10:30:19	1.1	31	3.41
10:30:39	1.1	31	3.41
10:30:59	1.1	31	3.41
10:31:19	1.1	31	3.41
10:31:39	1.2	31	4.01
10:31:59	1.1	31	3.41
10:32:19	1.1	31	3.41
10:32:39	1.1	31	3.41
10:32:59	1.1	31	3.41
10:33:19	1.1	31	3.41
10:33:39	1.2	31	4.01
10:33:59	1.2	31	4.01
10:34:19	1.2	31	4.01
10:34:39	1.1	31	3.41
10:34:59	1.1	31	3.41
10:35:19	1.2	31	4.01
10:35:39	1.2	31	4.01
10:35:59	1.1	31	3.41

Job Ref: OEH 31689
Client Name: LTI
Location: Combi Booth
Date: 19-Nov-03
Scientist: BR

Technical Details
Instrument Type: FID
Calib Gas: Methane
% C: 75
Sample Number: SA 5376
Instrument Range: 440
Emission Limit: 50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
10:36:19	1.1	31	3.41
10:36:39	1.1	31	3.41
10:36:59	1.1	31	3.41
10:37:19	1.1	31	3.41
10:37:39	1.1	31	3.41
10:37:59	1.1	31	3.41
10:38:19	1.6	31	5.15
10:38:39	1.1	31	3.41
10:38:59	1.4	31	4.54
10:39:19	1.1	31	3.41
10:39:39	1.1	31	3.41
10:39:59	0.9	31	2.88
10:40:19	0.9	31	2.88
10:40:39	0.9	31	2.88
10:40:59	0.9	31	2.88
10:41:19	0.9	31	2.88
10:41:39	0.9	31	2.88
10:41:59	0.9	31	2.88
10:42:19	0.9	31	2.88
10:42:39	0.9	31	2.88
10:42:59	0.9	31	2.88
10:43:19	0.9	31	2.88
10:43:39	0.9	31	2.88
10:43:59	0.9	31	2.88
10:44:19	0.9	31	2.88
10:44:39	0.9	31	2.88
10:44:59	0.9	31	2.88
10:45:19	1.2	31	4.01
10:45:39	1.2	31	4.01
10:45:59	1.1	31	3.41
10:46:19	1.1	31	3.41
10:46:39	1.1	31	3.41
10:46:59	1.1	31	3.41
10:47:19	1.1	31	3.41
10:47:39	0.9	31	2.88
10:47:59	1.1	31	3.41
10:48:19	1.1	31	3.41
10:48:39	0.9	31	2.88
10:48:59	1.1	31	3.41
10:49:19	1.2	31	4.01
10:49:39	1.1	31	3.41
10:49:59	1.1	31	3.41
10:50:19	1.1	31	3.41
10:50:39	1.1	31	3.41

Job Ref: OEH 31689
Client Name: LTI
Location: Combi Booth
Date: 19-Nov-03
Scientist: BR

Technical Details
Instrument Type: FID
Calib Gas: Methane
% C: 75
Sample Number: SA 5376
Instrument Range: 440
Emission Limit: 50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
10:50:59	1.1	31	3.41
10:51:19	1.1	31	3.41
10:51:39	1.1	31	3.41
10:51:59	1.1	31	3.41
10:52:19	1.1	31	3.41
10:52:39	1.2	31	4.01
10:52:59	1.4	31	4.54
10:53:19	1.4	31	4.54
10:53:39	1.2	31	4.01
10:53:59	1.2	31	4.01
10:54:19	1.2	31	4.01
10:54:39	1.2	31	4.01
10:54:59	1.2	31	4.01
10:55:19	1.2	31	4.01
10:55:39	1.2	31	4.01
10:55:59	1.8	31	5.68
10:56:19	1.4	31	4.54
10:56:39	1.4	31	4.54
10:56:59	1.4	31	4.54
10:57:19	1.2	31	4.01
10:57:39	1.2	31	4.01
10:57:59	1.4	31	4.54
10:58:19	1.4	31	4.54
10:58:39	1.4	31	4.54
10:58:59	1.6	31	5.15
10:59:19	2.0	31	6.29
10:59:39	1.8	31	5.68
10:59:59	1.6	31	5.15
11:00:19	2.0	31	6.29
11:00:39	2.0	31	6.29
11:00:59	1.8	31	5.68
11:01:19	1.8	31	5.68
11:01:39	2.1	31	6.82
11:01:59	1.6	31	5.15
11:02:19	1.8	31	5.68
11:02:39	2.0	31	6.29
11:02:59	1.8	31	5.68
11:03:19	1.6	31	5.15
11:03:39	1.6	31	5.15
11:03:59	1.6	31	5.15
11:04:19	1.6	31	5.15
11:04:39	1.8	31	5.68
11:04:59	1.4	31	4.54
11:05:19	1.4	31	4.54

Job Ref:	OEH 31689	Technical Details	
Client Name:	LTI	Instrument Type	FID
Location:	Combi Booth	Calib Gas	Methane
Date:	19-Nov-03	% C:	75
Scientist:	BR	Sample Number:	SA 5376
		Instrument Range:	440
		Emission Limit:	50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
11:05:39	1.4	31	4.54
11:05:59	1.4	31	4.54
11:06:19	1.4	31	4.54
11:06:39	1.4	31	4.54
11:06:59	1.4	31	4.54
11:07:19	1.4	31	4.54
11:07:39	1.4	31	4.54
11:07:59	1.4	31	4.54
11:08:19	1.2	31	4.01
11:08:39	1.2	31	4.01
11:08:59	1.2	31	4.01
11:09:19	1.4	31	4.54
11:09:39	1.4	31	4.54
11:09:59	1.2	31	4.01
11:10:19	1.4	31	4.54
11:10:39	1.4	31	4.54
11:10:59	1.4	31	4.54
11:11:19	1.2	31	4.01
11:11:39	1.8	31	5.68
11:11:59	1.4	31	4.54
11:12:19	1.2	31	4.01
11:12:39	1.2	31	4.01
11:12:59	1.4	31	4.54
11:13:19	1.4	31	4.54
11:13:39	1.2	31	4.01
11:13:59	1.2	31	4.01
11:14:19	1.2	31	4.01
11:14:39	1.2	31	4.01
11:14:59	1.4	31	4.54
11:15:19	1.4	31	4.54
11:15:39	1.4	31	4.54
11:15:59	1.2	31	4.01
11:16:19	1.2	31	4.01
11:16:39	1.2	31	4.01
11:16:59	1.2	31	4.01
11:17:19	1.2	31	4.01
11:17:39	1.4	31	4.54
11:17:59	1.2	31	4.01
11:18:19	1.2	31	4.01
11:18:39	1.2	31	4.01
11:18:59	1.2	31	4.01
11:19:19	1.2	31	4.01
11:19:39	1.2	31	4.01
11:19:59	1.2	31	4.01

VOC Emission Data

Job Ref:

OEH 31689

Technical Details

Client Name:

LTI

Instrument Type

FID

Location:

Combi Booth

Calib Gas

Methane

Date:

19-Nov-03

% C:

75

Scientist:

BR

Sample Number:

SA 5376

Instrument Range:

440

Emission Limit:

50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
11:20:19	1.2	31	4.01
11:20:39	1.2	31	4.01
11:20:59	1.2	31	4.01
11:21:19	1.2	31	4.01
11:21:39	1.2	31	4.01
11:21:59	1.2	31	4.01
11:22:19	1.1	31	3.41
11:22:39	1.1	31	3.41
11:22:59	1.2	31	4.01
11:23:19	1.2	31	4.01
11:23:39	1.2	31	4.01
11:23:59	1.4	31	4.54
11:24:19	1.2	31	4.01
11:24:39	1.2	31	4.01
11:24:59	1.2	31	4.01
11:25:19	1.1	31	3.41
11:25:39	1.2	31	4.01
11:25:59	1.6	31	5.15
11:26:19	1.2	31	4.01
11:26:39	1.2	31	4.01
11:26:59	1.2	31	4.01
11:27:19	1.4	31	4.54
11:27:39	2.1	31	6.82
11:27:59	3.0	31	9.69
11:28:19	3.7	31	11.97
11:28:39	3.0	31	9.69
11:28:59	1.8	31	5.68
11:29:19	3.0	31	9.69
11:29:39	2.8	31	9.09
11:29:59	4.3	31	13.71
11:30:19	3.2	31	10.22
11:30:39	3.0	31	9.69
11:30:59	2.0	31	6.29
11:31:19	2.8	31	9.09
11:31:39	3.0	31	9.69
11:31:59	2.7	31	8.56
11:32:19	2.7	31	8.56
11:32:39	2.3	31	7.42
11:32:59	2.1	31	6.82
11:33:19	2.1	31	6.82
11:33:39	2.1	31	6.82
11:33:59	2.0	31	6.29
11:34:19	2.1	31	6.82
11:34:39	2.1	31	6.82

Job Ref:

OEH 31689

Technical Details

Client Name:

LTI

Instrument Type

FID

Location:

Combi Booth

Calib Gas

Methane

Date:

19-Nov-03

% C:

75

Scientist:

BR

Sample Number:

SA 5376

Instrument Range:

440

Emission Limit:

50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
11:34:59	2.5	31	7.95
11:35:19	2.5	31	7.95
11:35:39	2.7	31	8.56
11:35:59	2.8	31	9.09
11:36:19	2.1	31	6.82
11:36:39	2.1	31	6.82
11:36:59	2.1	31	6.82
11:37:19	2.1	31	6.82
11:37:39	4.1	31	13.10
11:37:59	2.1	31	6.82
11:38:19	2.0	31	6.29
11:38:39	1.8	31	5.68
11:38:59	1.8	31	5.68
11:39:19	1.8	31	5.68
11:39:39	1.8	31	5.68
11:39:59	1.8	31	5.68
11:40:19	2.1	31	6.82
11:40:39	2.1	31	6.82
11:40:59	2.0	31	6.29
11:41:19	1.8	31	5.68
11:41:39	1.6	31	5.15
11:41:59	1.6	31	5.15
11:42:19	1.6	31	5.15
11:42:39	1.4	31	4.54
11:42:59	1.4	31	4.54
11:43:19	2.5	31	7.95
11:43:39	1.4	31	4.54
11:43:59	1.6	31	5.15
11:44:19	1.4	31	4.54
11:44:39	1.4	31	4.54
11:44:59	1.6	31	5.15
11:45:19	1.8	31	5.68
11:45:39	2.3	31	7.42
11:45:59	7.4	31	23.93
11:46:19	9.1	31	29.08
11:46:39	3.0	31	9.69
11:46:59	9.1	31	29.08
11:47:19	18.8	31	60.44
11:47:39	13.7	31	43.93
11:47:59	19.0	31	60.97
11:48:19	8.9	31	28.48
11:48:39	11.0	31	35.37
11:48:59	22.0	31	70.66
11:49:19	20.4	31	65.59

Job Ref:

OEH 31689

Technical Details

Client Name:

LTI

Instrument Type

FID

Location:

Combi Booth

Calib Gas

Methane

Date:

19-Nov-03

% C:

75

Scientist:

BR

Sample Number:

SA 5376

Instrument Range:

440

Emission Limit:

50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
11:49:39	23.2	31	74.67
11:49:59	10.3	31	33.10
11:50:19	6.0	31	19.39
11:50:39	17.6	31	56.42
11:50:59	13.5	31	43.32
11:51:19	6.9	31	22.27
11:51:39	5.0	31	15.98
11:51:59	10.5	31	33.63
11:52:19	3.0	31	9.69
11:52:39	2.5	31	7.95
11:52:59	2.5	31	7.95
11:53:19	2.3	31	7.42
11:53:39	2.1	31	6.82
11:53:59	2.1	31	6.82
11:54:19	2.0	31	6.29
11:54:39	2.0	31	6.29
11:54:59	2.0	31	6.29
11:55:19	2.5	31	7.95
11:55:39	2.3	31	7.42
11:55:59	2.3	31	7.42
11:56:19	2.5	31	7.95
11:56:39	2.5	31	7.95
11:56:59	2.5	31	7.95
11:57:19	2.3	31	7.42
11:57:39	2.3	31	7.42
11:57:59	2.7	31	8.56
11:58:19	2.5	31	7.95
11:58:39	2.3	31	7.42
11:58:59	2.1	31	6.82
11:59:19	2.1	31	6.82
11:59:39	2.1	31	6.82
11:59:59	2.3	31	7.42
12:00:19	2.3	31	7.42
12:00:39	2.1	31	6.82
12:00:59	2.0	31	6.29
12:01:19	2.0	31	6.29
12:01:39	1.8	31	5.68
12:01:59	1.6	31	5.15
12:02:19	1.6	31	5.15
12:02:39	1.8	31	5.68
12:02:59	1.8	31	5.68
12:03:19	1.8	31	5.68
12:03:39	1.6	31	5.15
12:03:59	1.8	31	5.68

Job Ref:	OEH 31689	Technical Details	
Client Name:	LTI	Instrument Type	FID
Location:	Combi Booth	Calib Gas	Methane
Date:	19-Nov-03	% C:	75
Scientist:	BR	Sample Number:	SA 5376
		Instrument Range:	440
		Emission Limit:	50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
12:04:19	1.8	31	5.68
12:04:39	1.6	31	5.15
12:04:59	1.6	31	5.15
12:05:19	1.4	31	4.54
12:05:39	1.6	31	5.15
12:05:59	1.6	31	5.15
12:06:19	1.6	31	5.15
12:06:39	1.6	31	5.15
12:06:59	1.4	31	4.54
12:07:19	1.6	31	5.15
12:07:39	1.6	31	5.15
12:07:59	1.6	31	5.15
12:08:19	1.8	31	5.68
12:08:39	1.8	31	5.68
12:08:59	1.4	31	4.54
12:09:19	1.4	31	4.54
12:09:39	1.4	31	4.54
12:09:59	1.2	31	4.01
12:10:19	1.2	31	4.01
12:10:39	1.4	31	4.54
12:10:59	1.4	31	4.54
12:11:19	1.4	31	4.54
12:11:39	1.4	31	4.54
12:11:59	1.2	31	4.01
12:12:19	1.4	31	4.54
12:12:39	1.4	31	4.54
12:12:59	1.4	31	4.54
12:13:19	1.2	31	4.01
12:13:39	1.2	31	4.01
12:13:59	1.2	31	4.01
12:14:19	1.2	31	4.01
12:14:39	1.4	31	4.54
12:14:59	1.6	31	5.15
12:15:19	1.8	31	5.68
12:15:39	1.8	31	5.68
12:15:59	1.4	31	4.54
12:16:19	1.4	31	4.54
12:16:39	1.4	31	4.54
12:16:59	1.4	31	4.54
12:17:19	1.2	31	4.01
12:17:39	1.4	31	4.54
12:17:59	1.2	31	4.01
12:18:19	1.2	31	4.01
12:18:39	1.4	31	4.54

VOC Emission Data

Job Ref:

OEH 31689

Technical Details

Client Name:

LTI

Instrument Type

FID

Location:

Combi Booth

Calib Gas

Methane

Date:

19-Nov-03

% C:

75

Scientist:

BR

Sample Number:

SA 5376

Instrument Range:

440

Emission Limit:

50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
12:18:59	1.2	31	4.01
12:19:19	1.6	31	5.15
12:19:39	1.4	31	4.54
12:19:59	1.2	31	4.01
12:20:19	1.2	31	4.01
12:20:39	1.2	31	4.01
12:20:59	1.2	31	4.01
12:21:19	1.2	31	4.01
12:21:39	1.2	31	4.01
12:21:59	1.2	31	4.01
12:22:19	1.2	31	4.01
12:22:39	1.1	31	3.41
12:22:59	1.1	31	3.41
12:23:19	1.1	31	3.41
12:23:39	1.1	31	3.41
12:23:59	1.1	31	3.41
12:24:19	1.1	31	3.41
12:24:39	1.1	31	3.41
12:24:59	1.1	31	3.41
12:25:19	1.1	31	3.41
12:25:39	1.1	31	3.41
12:25:59	1.1	31	3.41
12:26:19	1.1	31	3.41
12:26:39	1.1	31	3.41
12:26:59	1.1	31	3.41
12:27:19	1.1	31	3.41
12:27:39	1.1	31	3.41
12:27:59	1.1	31	3.41
12:28:19	1.1	31	3.41
12:28:39	1.1	31	3.41
12:28:59	0.9	31	2.88
12:29:19	0.9	31	2.88
12:29:39	1.1	31	3.41
12:29:59	1.1	31	3.41
12:30:19	1.1	31	3.41
12:30:39	0.9	31	2.88
12:30:59	0.9	31	2.88
12:31:19	1.1	31	3.41
12:31:39	1.1	31	3.41
12:31:59	0.9	31	2.88
12:32:19	0.9	31	2.88
12:32:39	0.9	31	2.88
12:32:59	1.1	31	3.41
12:33:19	0.9	31	2.88

Job Ref:	OEH 31689	Technical Details	
Client Name:	LTI	Instrument Type	FID
Location:	Combi Booth	Calib Gas	Methane
Date:	19-Nov-03	% C:	75
Scientist:	BR	Sample Number:	SA 5376
		Instrument Range:	440
		Emission Limit:	50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
12:33:39	0.9	31	2.88
12:33:59	0.9	31	2.88
12:34:19	0.9	31	2.88
12:34:39	0.9	31	2.88
12:34:59	0.9	31	2.88
12:35:19	0.9	31	2.88
12:35:39	0.9	31	2.88
12:35:59	0.9	31	2.88
12:36:19	0.9	31	2.88
12:36:39	0.9	31	2.88
12:36:59	0.9	31	2.88
12:37:19	0.9	31	2.88
12:37:39	0.9	31	2.88
12:37:59	0.7	31	2.27
12:38:19	0.9	31	2.88
12:38:39	0.9	31	2.88
12:38:59	0.9	31	2.88
12:39:19	0.9	31	2.88
12:39:39	0.9	31	2.88
12:39:59	0.9	31	2.88
12:40:19	0.7	31	2.27
12:40:39	0.9	31	2.88
12:40:59	0.9	31	2.88
12:41:19	0.9	31	2.88
12:41:39	0.9	31	2.88
12:41:59	0.9	31	2.88
12:42:19	0.9	31	2.88
12:42:39	0.9	31	2.88
12:42:59	0.9	31	2.88
12:43:19	0.7	31	2.27
12:43:39	0.9	31	2.88
12:43:59	0.9	31	2.88
12:44:19	0.9	31	2.88
12:44:39	0.9	31	2.88
12:44:59	0.7	31	2.27
12:45:19	0.9	31	2.88
12:45:39	0.7	31	2.27
12:45:59	0.7	31	2.27
12:46:19	0.7	31	2.27
12:46:39	0.7	31	2.27
12:46:59	0.7	31	2.27
12:47:19	0.9	31	2.88
12:47:39	0.7	31	2.27
12:47:59	0.7	31	2.27

VOC Emission Data

Job Ref:

OEH 31689

Technical Details

Client Name:

LTI

Instrument Type

FID

Location:

Combi Booth

Calib Gas

Methane

Date:

19-Nov-03

% C:

75

Scientist:

BR

Sample Number:

SA 5376

Instrument Range:

440

Emission Limit:

50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
12:48:19	0.7	31	2.27
12:48:39	0.7	31	2.27
12:48:59	0.7	31	2.27
12:49:19	0.7	31	2.27
12:49:39	0.7	31	2.27
12:49:59	0.7	31	2.27
12:50:19	0.7	31	2.27
12:50:39	0.7	31	2.27
12:50:59	0.7	31	2.27
12:51:19	0.7	31	2.27
12:51:39	0.7	31	2.27
12:51:59	0.9	31	2.88
12:52:19	0.7	31	2.27
12:52:39	0.7	31	2.27
12:52:59	0.7	31	2.27
12:53:19	0.7	31	2.27
12:53:39	0.7	31	2.27
12:53:59	0.7	31	2.27
12:54:19	0.7	31	2.27
12:54:39	0.7	31	2.27
12:54:59	0.7	31	2.27
12:55:19	0.7	31	2.27
12:55:39	0.7	31	2.27
12:55:59	0.7	31	2.27
12:56:19	0.7	31	2.27
12:56:39	0.7	31	2.27
12:56:59	0.7	31	2.27
12:57:19	0.7	31	2.27
12:57:39	0.7	31	2.27
12:57:59	0.7	31	2.27
12:58:19	0.9	31	2.88
12:58:39	0.7	31	2.27
12:58:59	0.7	31	2.27
12:59:19	0.7	31	2.27
12:59:39	0.7	31	2.27
12:59:59	0.7	31	2.27
13:00:19	0.7	31	2.27
13:00:39	0.7	31	2.27
13:00:59	0.7	31	2.27
13:01:19	0.7	31	2.27
13:01:39	0.7	31	2.27
13:01:59	0.7	31	2.27
13:02:19	1.2	31	4.01
13:02:39	1.1	31	3.41

VOC Emission Data

Job Ref:	OEH 31689	Technical Details	
Client Name:	LTI	Instrument Type	FID
Location:	Combi Booth	Calib Gas	Methane
Date:	19-Nov-03	% C:	75
Scientist:	BR	Sample Number:	SA 5376
		Instrument Range:	440
		Emission Limit:	50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
13:02:59	1.1	31	3.41
13:03:19	0.9	31	2.88
13:03:39	1.1	31	3.41
13:03:59	0.9	31	2.88
13:04:19	0.9	31	2.88
13:04:39	0.9	31	2.88
13:04:59	0.9	31	2.88
13:05:19	0.7	31	2.27
13:05:39	0.9	31	2.88
13:05:59	0.9	31	2.88
13:06:19	0.7	31	2.27
13:06:39	0.9	31	2.88
13:06:59	0.9	31	2.88
13:07:19	0.9	31	2.88
13:07:39	0.7	31	2.27
13:07:59	0.7	31	2.27
13:08:19	0.9	31	2.88
13:08:39	0.9	31	2.88
13:08:59	0.9	31	2.88
13:09:19	0.9	31	2.88
13:09:39	0.9	31	2.88
13:09:59	0.9	31	2.88
13:10:19	0.9	31	2.88
13:10:39	0.9	31	2.88
13:10:59	1.1	31	3.41
13:11:19	0.9	31	2.88
13:11:39	0.9	31	2.88
13:11:59	0.9	31	2.88
13:12:19	0.9	31	2.88
13:12:39	0.9	31	2.88
13:12:59	1.1	31	3.41
13:13:19	0.9	31	2.88
13:13:39	0.9	31	2.88
13:13:59	0.9	31	2.88
13:14:19	0.9	31	2.88
13:14:39	0.9	31	2.88
13:14:59	0.9	31	2.88
13:15:19	0.9	31	2.88
13:15:39	0.9	31	2.88
13:15:59	0.9	31	2.88
13:16:19	0.9	31	2.88
13:16:39	1.1	31	3.41
13:16:59	1.1	31	3.41
13:17:19	1.1	31	3.41

Job Ref:	OEH 31689	Technical Details	
Client Name:	LTI	Instrument Type	FID
Location:	Combi Booth	Calib Gas	Methane
Date:	19-Nov-03	% C:	75
Scientist:	BR	Sample Number:	SA 5376
		Instrument Range:	440
		Emission Limit:	50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
13:17:39	1.1	31	3.41
13:17:59	1.1	31	3.41
13:18:19	1.1	31	3.41
13:18:39	1.1	31	3.41
13:18:59	1.1	31	3.41
13:19:19	1.1	31	3.41
13:19:39	1.1	31	3.41
13:19:59	0.9	31	2.88
13:20:19	0.9	31	2.88
13:20:39	0.9	31	2.88
13:20:59	0.9	31	2.88
13:21:19	0.9	31	2.88
13:21:39	0.9	31	2.88
13:21:59	0.9	31	2.88
13:22:19	1.1	31	3.41
13:22:39	1.1	31	3.41
13:22:59	1.1	31	3.41
13:23:19	0.9	31	2.88
13:23:39	0.9	31	2.88
13:23:59	0.9	31	2.88
13:24:19	0.9	31	2.88
13:24:39	0.9	31	2.88
13:24:59	0.9	31	2.88
13:25:19	0.9	31	2.88
13:25:39	0.9	31	2.88
13:25:59	0.9	31	2.88
13:26:19	0.9	31	2.88
13:26:39	0.9	31	2.88
13:26:59	0.9	31	2.88
13:27:19	0.9	31	2.88
13:27:39	0.9	31	2.88
13:27:59	0.9	31	2.88
13:28:19	0.9	31	2.88
13:28:39	0.9	31	2.88
13:28:59	0.9	31	2.88
13:29:19	0.9	31	2.88
13:29:39	0.9	31	2.88
13:29:59	0.9	31	2.88
13:30:19	1.1	31	3.41
13:30:39	0.9	31	2.88
13:30:59	0.9	31	2.88
13:31:19	1.1	31	3.41
13:31:39	1.1	31	3.41
13:31:59	1.1	31	3.41

VOC Emission Data

Job Ref:

OEH 31689

Technical Details

Client Name:

LTI

Instrument Type

FID

Location:

Combi Booth

Calib Gas

Methane

Date:

19-Nov-03

% C:

75

Scientist:

BR

Sample Number:

SA 5376

Instrument Range:

440

Emission Limit:

50

Sampling Time	VOC as Methane Equivalent@ Reference Conditions (mgC.m ⁻³)	Stack Temp (°C)	VOC as Carbon @ Reference Conditions (Solvent Response Corrected) (mgC.m ⁻³)
13:32:19	1.1	31	3.41
13:32:39	1.1	31	3.41
13:32:59	1.2	31	4.01
13:33:19	1.2	31	4.01
13:33:39	1.1	31	3.41
13:33:59	1.1	31	3.41
13:34:19	1.1	31	3.41
13:34:39	1.1	31	3.41
13:34:59	1.1	31	3.41
13:35:19	1.1	31	3.41
13:35:39	1.1	31	3.41
13:35:59	1.1	31	3.41
13:36:19	1.1	31	3.41
13:36:39	1.1	31	3.41
13:36:59	1.1	31	3.41
13:37:19	1.1	31	3.41
13:37:39	1.1	31	3.41
13:37:59	1.1	31	3.41
13:38:19	1.1	31	3.41
13:38:39	1.1	31	3.41
13:38:59	1.1	31	3.41
13:39:19	1.1	31	3.41
13:39:39	1.1	31	3.41
13:39:59	1.1	31	3.41
13:40:19	1.1	31	3.41
13:40:39	1.1	31	3.41
13:40:59	1.1	31	3.41
13:41:19	1.1	31	3.41
13:41:39	1.1	31	3.41
13:41:59	1.1	31	3.41
13:42:19	1.1	31	3.41
13:42:39	1.1	31	3.41
13:42:59	1.1	31	3.41
13:43:19	1.1	31	3.41
13:43:39	1.1	31	3.41
13:43:59	1.1	31	3.41
13:44:19	1.1	31	3.41
13:44:39	1.1	31	3.41
13:44:59	1.1	31	3.41
13:45:19	1.1	31	3.41
Average	1.5	Average	4.9

**APPENDIX III
CALIBRATION CERTIFICATES**

LEVEL 3: WORK INSTRUCTIONS – CONSULTANCY

CSOP ??: METHODS FOR EQUIPMENT CALIBRATION – STACKMITE GAS METERS AND ROTAMETERS

STACKMITE CALIBRATION DATA SHEET

Name: ANDY BARNES
 Signed: [REDACTED]
 Date: 31/03/03
 Test Stackmite Ref: A0 001
 Reference Gas Meter Ref: A0 046
 Reference Gas Meter Calibration Date: 04/03/03
 Reference Gas Meter Calibration Certificate No: N 1096 766F

Section 1 – Test Readings

Test Meter Readings (litres)			Time		Reference Meter Readings (litres)			Indicated Flow on Rotameter (l/min)
Start	Stop	Diff	Start	Stop	Start	Stop	Diff	
173635	174217	582	1141	1211	5101	5692	591	20
174220	174670	450	1212	1242	5695	6146	451	15
174675	174979	304	1243	1313	6150	6452	302	10
174985	175139	154	1315	1345	6457	6608	151	5

Section 2 – Calculations

Indicated Flow on Rotameter (l/min)	Actual Flowrate (l/min)	Rotameter % Error	Test Meter Volume (litres)	Reference Meter Volume (litres)	Gas Meter % Error
20	19.70	-1.50	582	591	-1.52
15	15.03	+0.22	450	451	-0.22
10	10.07	+0.70	304	302	+0.66
5	5.03	+0.60	154	151	+1.99%

OEH GROUP LIMITED
AS500 LOW FLOW PUMP - CERTIFICATE OF CONFORMITY

Pump No.	Run Time (min)	Start Count	Stop Count	Pump Setting (0-100)	Measured flow rate (ml/min (average of 5 readings over period))	Calculated volume per stroke (ml/stroke)	Initial	Signature	Date
AQ029	20	481404	484367	20	70.56	0.48	WR	W Roberts	19 6 03
AQ029	21	484367	489453	40	116.2	0.48	WR	W Roberts	19 6 03
						#DIV/0!			
						#DIV/0!			
						#DIV/0!			
						#DIV/0!			
						#DIV/0!			

SORBSAT INCE / ISO CYANATE SAMPLING PUMP

OEH GROUP LIMITED
AS500 LOW FLOW PUMP - CERTIFICATE OF CONFORMITY

Pump No.	Run Time (min)	Start Count	Stop Count	Pump Setting (0.5-100)	Measured flow rate (ml/min (average of 5 readings over period))	Calculated volume per stroke (ml/stroke)	Initial	Signature	Date
AQ021	20	171414	176195	40	110	0.46	WR	WRoberts	18603
AQ021	20	176195	179292	20	70.5	0.46	WR	WRoberts	18603
						#DIV/0!			
						#DIV/0!			
						#DIV/0!			
						#DIV/0!			
						#DIV/0!			

SORBENT TRAY / ISO CYANURATE SAMPLING PUMP.

A0046

CERTIFICATE OF CALIBRATION



DATE OF ISSUE 04 March 2003

CERTIFICATE NUMBER N1096766F

0072

Wyko Calibration Services
 Romsey Laboratory
 UNIVERSAL HOUSE
 ROMSEY INDUSTRIAL ESTATE
 ROMSEY
 HAMPSHIRE
 SO51 0HR
 Telephone 01794 523935
 Facsimile 01794 523910
 Email cal.services@wyko.co.uk
 Website www.cal.wyko.co.uk



...much more than you imagine

PAGE 1 OF 2 PAGES

APPROVED SIGNATORY

[] P. Davies
 [] W. Hatt
 []
 []

Customer : OEH GROUP LTD
 Address : 253-255 GREAT LISTER STREET
 BIRMINGHAM
 B7 4BS

Order No : 3564

Apparatus Tested - DRY GAS METER AQ 046 REFERENCE METAL.

Type No : SK25

Serial No : 0003729

Inventory No : -

Manufacturer : KIMMON MFG CO LTD

Range/Scale : Calibrated range: 0.5 TO 2.5 m³/h
 Resolution: 0.001 m³/h, QMax 2.5 m³/h


Test Conditions -

Date Instrument Received : 03 March 2003

Date Calibration Completed : 04 March 2003

Ambient Temperature : 20 ± 2 °C

Reference No : 1096766

Certified by 

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

LEVEL 3: WORK INSTRUCTIONS – CONSULTANCY

CSOP ??: METHODS FOR EQUIPMENT CALIBRATION – TAPE MEASURES FOR STACK CONSULTANCY

MEASURING TAPE CALIBRATION DATA SHEET

Name: W Roberts
Signed: W Roberts
Date: 17 March 03
Test Measuring Tape Ref: AO 047
Test Measuring Tape Length: 5m
Reference Ruler Ref: AO 045
Reference Ruler Calibration Date: 28 Feb 03
Reference Ruler Calibration Certificate No: 98275

Test Tape Interval	Reference Ruler Reading	Error (%)	Allowed Tolerance	Pass/Fail?
0.000 – 1.000m	1.000	—	± 0.5%	✓
1.000 – 2.000m	1.000	—	± 0.5%	✓
2.000 – 3.000m	1.000	—	± 0.5%	✓
3.000 – 4.000m	1.000	—	± 0.5%	✓
4.000 – 5.000m	1.000	—	± 0.5%	✓
5.000 – 6.000m			± 0.5%	
6.000 – 7.000m			± 0.5%	

CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date of Issue: 28-Feb-2003

Certificate No. 98275



0310

RS Calibration

Calibration and Repair Service

Venture Close, Lammas Rd, Corby,
Northants NN17 5UB

Tel: 01536 405545

Fax: 01536 401590

Page 1 of 3 pages

Approved signatory


K. O'Malley

CUSTOMER	OEH Group Ltd. Birmingham B7 4BS
INSTRUMENT	1m/39in Steel Rule.
PROCEDURE ID.	MLCP57
MANUFACTURER	Rabone
SERIAL No.	RS12974
DATE RECEIVED	N/A
DATE CALIBRATED	28-Feb-2003

AG045

REFERENCE QUV

REMARKS

The measured results were found to be within the tolerances specified in BS 4372:1968.

Uncertainties

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories.
This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

DEUTSCHER KALIBRIERDIENST

DKD

A0064

Kalibrierlaboratorium für die Meßgröße Druck
Calibration laboratory for pressure measurement

MANOMETER

Akkreditiert durch die / accredited by the

Akkreditierungsstelle des DKD bei der
PHYSIKALISCH-TECHNISCHEN BUNDESANSTALT (PTB)



DKD-K-11204



industrial services

calibration validation training

Kalibrierschein

Calibration Certificate
Certificat d'étalonnage

Kalibrierzeichen

Calibration mark
Marque d'étalonnage

0648
DKD-K-11204
03-08

Gegenstand Druckmessgerät pressure instrument

Object

Objet

Hersteller

Manufacturer

Fabricant

testo AG
D-79853 Lenzkirch

Typ

Type

Type

0560 5121

Fabrikat/Serien-Nr.

Serial number

Numéro de série

30208034

Auftraggeber

Customer

Client

E.I.M. (Northern) Ltd.
GB-SK1 3HZ Stockport

Auftragsnummer

Order No.

Numéro de commande

185546

Anzahl der Seiten des Kalibrierscheines - 5 -

Number of pages of the certificate

Numéro de pages

Datum der Kalibrierung

Date of calibration

Date d'étalonnage

15.08.2003

Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI).

Der DKD ist Unterzeichner der multilateralen Übereinkommen der European co-operation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur gegenseitigen Anerkennung der Kalibrierscheine.

Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.

This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI).

The DKD is signatory to the multilateral agreements of the European co-operation for Accreditation (EA) and of the International Laboratory Accreditation Cooperation (ILAC) for the mutual recognition of calibration certificates.

The user is obliged to have the object recalibrated at appropriate intervals.

Ce certificat d'étalonnage documente la traçabilité des grandeurs mesurées par raccordement aux étalons nationaux en conformité avec le Système international d'unités (SI).

Le DKD est signataire des accords multi-latéraux de la European co-operation for Accreditation (EA) et de la International Laboratory Accreditation Cooperation (ILAC) pour la reconnaissance mutuelle des certificats d'étalonnage.

L'utilisateur est tenu de faire étalonner le matériel référencé ci-dessus à des intervalles appropriés.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of both the Accreditation Body of the DKD and the issuing laboratory. Calibration certificates without signature and seal are not valid.

Ce Certificat d'étalonnage ne doit être divulgué que dans sa forme complète et sans modifications. Des extraits ou modifications doivent être autorisés par le Service d'accréditation du DKD et par le laboratoire d'étalonnage ayant établi le certificat. Les certificats d'étalonnage non signés et non estampillés ne sont pas valides.



Datum

Date

Date

15.08.2003

Stellv. Leiter des Kalibrierlaboratoriums

Head of the calibration laboratory

Directeur du laboratoire d'étalonnage

Falko Harich

Bearbeiter

Person in charge

Personne responsable

Melanie Mink

Seite 2 zum Kalibrierschein vom 15.08.2003

Kalibriergegenstand:

Typ: Druckmeßgerät
Messart: positiver Überdruck
Messbereich: 0...20 mbar
Auflösung: 0,01 mbar
Genauigkeit: $\pm 0,5\%$ v.Ew.

Bezugsnormal:

Modell: V1600/1, Ser.-Nr.:11447-99
Messbereich 0,2-160 mbar
Messunsicherheit: $2 \cdot 10^{-4} \cdot p_e$, jedoch nicht kleiner als 1 μ bar
Kalibrierzeichen: 0072 PTB 99

Umgebungsbedingungen:

Temperatur: (20 ± 3) °C
Luftdruck: (930 ± 30) mbar
Fallbeschleunigung: $(9,80681 \pm 2 \cdot 10^{-5}) \text{ m/s}^2$

Messbedingungen:

Druckübertragungsmittel: Luft
Druckbezugsebene: Mitte Druckanschluß
Endwertbelastung: 2 min
Anzahl der Meßwerte: 34

Einbaulage: horizontal

Kalibrierverfahren:

Die Kalibrierung erfolgte nach DKD R6-1 (Dezember 1998) Ablauf B

Bemerkungen:

Messergebnisse:

pe in mbar	Ablesung am KG				Mittelwert in mbar	Mess- abweichung in mbar	Wiederhol- barkeit in mbar	Hysterese in mbar	U mit k=2 in mbar
	M1 auf in mbar	M2 ab in mbar	M3 auf in mbar	M4 ab in mbar					
0,000	0,00	0,02	0,00	0,01	0,008	0,008	0,000	0,020	0,017
2,006	1,97	1,99	1,98		1,983	-0,023	0,010	0,020	0,018
4,010	3,95	3,96	3,96		3,958	-0,052	0,010	0,010	0,015
6,010	5,93	5,96	5,94		5,948	-0,062	0,010	0,030	0,022
8,010	7,93	7,95	7,94		7,943	-0,067	0,010	0,020	0,018
10,010	9,95	9,96	9,96		9,958	-0,053	0,010	0,010	0,015
12,010	11,96	11,97	11,96		11,965	-0,045	0,000	0,010	0,014
14,009	13,92	13,92	13,94		13,925	-0,084	0,020	0,000	0,018
16,010	15,90	15,91	15,92		15,910	-0,100	0,020	0,010	0,019
18,009	17,90	17,91	17,92		17,910	-0,099	0,020	0,010	0,019
20,008	19,92	19,93	19,95		19,933	-0,075	0,030	0,010	0,023

Messunsicherheit:

Erst nach Korrektur des abgelesenen Druckwertes am Kalibriergegenstand mit dem Wert der Messabweichung (siehe Tabelle), entspricht der Betrag der in der Tabelle angegebenen Messunsicherheit.

maximale Messabweichung: 0,0998 mbar (0,499% vom Endwert)
 maximale Hysterese: 0,0300 mbar (0,150% vom Endwert)
 maximale Meßunsicherheit: 0,0228 mbar (0,114% vom Endwert)

Der Kalibriergegenstand hält die vom Hersteller festgelegten Fehlergrenzen ein.

Angegeben ist die erweiterte Messunsicherheit, die sich aus der Standardmessunsicherheit durch Multiplikation mit dem Erweiterungsfaktor $k=2$ ergibt. Sie wurde gemäß DKD-3 ermittelt. Der Wert der Messgröße liegt mit einer Wahrscheinlichkeit von 95% im zugeordneten Werteintervall.

Am Kalibriergegenstand ist eine Kalibriermarke angebracht, die mit der Kalibriernummer dieses DKD-Scheines, sowie mit dem Kalibriermonat und Jahr versehen wurde.

Der Deutsche Kalibrierdienst ist Unterzeichner des multilateralen Übereinkommens der European cooperation for Accreditation (EA) zur gegenseitigen Anerkennung der Kalibrierscheine. Die anderen Unterzeichner sind zur Zeit die Akkreditierungsstellen in Belgien, Dänemark, Finnland, Frankreich, Irland, Italien, den Niederlanden, Norwegen, Portugal, Österreich, Schweden, der Schweiz, Spanien, Tschechische Republik und dem Vereinigten Königreich. Ferner bestehen entsprechende Übereinkommen mit den Akkreditierungsstellen Australiens, Neuseelands und Südafrikas.



The German original text is valid in case of doubt.

Calibration object:

Type: Pressure measuring instrument
Measuring method: Positive overpressure
Measuring range : 0...20 mbar
Resolution: 0,01 mbar
Accuracy: $\pm 0,5\%$ v.Ew.

Reference standard:

Model: V1600/1, Ser.-Nr.:11447-99
Measuring range 0,2-160 mbar
Measuring uncertainty: $2 \cdot 10^{-4} \cdot p_e$, but not smaller than 1 μ bar
Calibration sign: 0072 PTB 99

Ambience condition:

Temperature: (20 ± 3) °C
Air pressure: (930 ± 30) mbar
acceleration of the fall: $(9,80681 \pm 2 \cdot 10^{-5})$ m/s²

Measurement conditions:

Pressure exchanger: air
Pressure level: Centre pressure connection
Limit value stress: 2 min
Number of measured values: 34

Fitting position: horizontal

Calibration method:

The calibration has been executed according to DKD R6-1 (December 1998) course B

Remarks:

The German original text is valid in case of doubt.

Measuring results:

Pressure mbar	Value				Average mbar	Deviation mbar	Repeatability mbar	Hysteresis mbar	U with k=2 mbar
	M1 up mbar	M2 down mbar	M3 up mbar	M4 down mbar					
0,000	0,00	0,02	0,00	0,01	0,008	0,008	0,000	0,020	0,017
2,006	1,97	1,99	1,98		1,983	-0,023	0,010	0,020	0,018
4,010	3,95	3,96	3,96		3,958	-0,052	0,010	0,010	0,015
6,010	5,93	5,96	5,94		5,948	-0,062	0,010	0,030	0,022
8,010	7,93	7,95	7,94		7,943	-0,067	0,010	0,020	0,018
10,010	9,95	9,96	9,96		9,958	-0,053	0,010	0,010	0,015
12,010	11,96	11,97	11,96		11,965	-0,045	0,000	0,010	0,014
14,009	13,92	13,92	13,94		13,925	-0,084	0,020	0,000	0,018
16,010	15,90	15,91	15,92		15,910	-0,100	0,020	0,010	0,019
18,009	17,90	17,91	17,92		17,910	-0,099	0,020	0,010	0,019
20,008	19,92	19,93	19,95		19,933	-0,075	0,030	0,010	0,023

Measuring uncertainty:

Only after correction of the pressure value reading on the calibration object with the value of the drift (see table), the amount equals the measuring uncertainty in the table.

Maximum drift:	0,0998	mbar	(0,499% of limit value)
Maximum hysteresis:	0,0300	mbar	(0,150% of limit value)
Maximum measuring uncertainty:	0,0228	mbar	(0,114% of limit value)

The calibration object meets the error limit determined by the manufacturer.

The extended measuring uncertainty, resulting of the standard measuring uncertainty multiplied with the extension factor k=2. It has been determined according to DKD-3. The value of the measurable is within the assigned value interval with a probability of 95%.

A calibration label has been applied to the calibration object, provided with the calibration number of this DKD-certificate as well as with the month and the year of the calibration.

The Deutsche Kalibrierdienst is signatory of the multilateral convention of the European cooperation for Accreditation (EA) for mutual acceptance of calibration certificates. The other signatories are at present the accreditation authorities in Belgium, Denmark, Finland, France, Ireland, Italy, Netherlands, Norway, Portugal, Austria, Sweden, Switzerland, Spain and the United Kingdom. Furthermore there are corresponding agreements with accreditation authorities in Australia, New Zealand and the Republic of South Africa.



AQ060

CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date of Issue: 15 Jul 2003

Certificate No. X13001



RS Calibration

Calibration and Repair Service

Venture Close, Lammas Rd, Corby,
Northants NN17 5UB

Tel: 01536 405545

Fax: 01536 401590

Approved Signatory: K.Mistry () S.Randall () M.Connelly () A.Bagshaw () J.Richards () I.Phillips () M.Tiney ()

Page 1 of 3 Pages

Client	OEH GROUP LTD. BIRMINGHAM B7 4BS
Instrument	Iso-Tech RS53 Temperature Indicator
Serial No.	23000112
Client Reference	N/A
Procedure ID.	206.3744 Wav9100 Rev P2
Date Received	N/A
Date of Calibration	15 Jul 2003

Remarks

This is a new instrument therefore no adjustment was necessary.

Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories.


This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF ANALYSIS

Date of Issue 8 th May 2003	Certificate No 030508-3-CC
--	----------------------------

Customer OEH Group Analyst N. Jones
Customer Order No 3605 Method of Analysis G. C.
Cylinder Type/Size 103 Litre Accuracy +/-2%
Cylinder No 2793 Std. Used 2793

Component	Required Concentration	Actual Concentration
Methane	90ppm	91.4ppm
Balance Air		

Analyst  Supervisor 

Traceability Statement

*All equipment used is traceable to NPL standards
{National Physics Laboratory Ref No. 086006 / MW 468-109}*

Bedfont
• S C I E N T I F I C • L T D •

105 Laker Road
Rochester Airport Ind.Est.
Rochester
Kent
ME1 3QX